



INSTRUMENTATION DETAILS			
SENSOR	DESCRIPTION	SETTING	ACTION
S1	BACKUP VAC SWT.	15" Hg ON THE FALL	BACKUP OPERATION
S2	VACUUM TRANSDUCER	VAC-15PSIG	OPERATION
S3	HIGH DISCH TEMP	5-10HP - 250°F	SHUTDOWN

—	AIR LINE
.....	DRAIN LINE
---	CUSTOMER PIPING
▷	PNEUMATIC FLOW
▶	CONDENSATE FLOW

Revised: EYP	Date: 11/01/2021	Drawn: JEB	Date: 12/01/10
This drawing and the information contained herein remain the property of BeaconMedas and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedas Engineering.			Scale: NTS
Description: PNEUMATIC DIAGRAM TX-X-QX 5-10HP LUBE VAC MOD		Part Number: 4107 8523 44	
DN: HOP211239		Rev: 01	
Sheet 1 of 1		DO NOT SCALE THIS DOCUMENT	