



Notes:

1. SYSTEM WEIGHT: 8,067 LBS.
(INCLUDES FUTURE EXPANSION PUMP)
2. INDICATES LIFTING POINT.
3. DENOTES CENTER OF GRAVITY.
(INCLUDES FUTURE EXPANSION PUMP)

SYSTEM CONNECTIONS	
ID	DESCRIPTION
I	INLET LOCATION
D	DISCHARGE LOCATION
C	CONDENSATE DRAIN LOCATION
T	TANK DRAIN LOCATION

MINIMUM CLEARANCE FOR
MAINTENANCE & VENTILATION

INTAKE ISOLATION
VALVE FOR FUTURE
EXPANSION PUMP

FUTURE
EXPANSION
PUMP

ELECTRICAL
CONNECTION

Ø0.75
BOLTDOWNS
(8 PLACES)

RECEIVER
BYPASS
VALVE

240 GALLON
RECEIVER

CONTROL CABINET

HEPA INTAKE FILTER
PER PUMP

DISCHARGE
3" FLANGE

BIOHAZARD
DRAIN FLASK
PER PUMP

INLET
4" FLANGE

INTEGRATED
LIFTING EYES

MANUAL DRAIN
1/2 " FNPT

MANUAL
CONDENSATE
DRAIN

Revised: a00545201	Date: 2/22/2023	Drawn: MDB	Date: 10/4/2010
		This drawing and the information contained therein remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering.	
Description: GENERAL ASSEMBLY 25HP DX-EXP-TX LUBE VAC SPC 240G RECEIVER		Part Number: 4107 8522 92	
Sheet 1 of 1		DN: HOP230756	Rev: 03
		DO NOT SCALE THIS DOCUMENT	