Notes: 1. DISCHARGE MUFFLER WILL BE SHIPPED LOOSE. 2. SYSTEM WEIGHT: 3,366 LBS. 24.00 3. INDICATES LIFTING POINT. **—24.00**— **--24.00**-4. DENOTES CENTER OF GRAVITY. SYSTEM CONNECTIONS DESCRIPTION MINIMUM CLEARANCE-INLET LOCATION FOR MAINTENANCE & VENTILATION 42.25 DISCHARGE LOCATION D TANK DRAIN LOCATION Ø0.75 BOLTDOWNS C CONDENSATE DRAIN LOCATION (8 PLACES) ELECTRICAL-CONNECTION -32.00--36.00 36.00 30.50 240 GALLON-**RECEIVER** -CONTROL CABINET HEPA INTAKE FILTER PER PUMP -DISCHARGE RECEIVER BYPASS-3" FLANGE **VALVE** 96.34 INLET-3" FLANGE D 54.94 33.17 I 23.87 BIOHAZARD C 1.97 T8.00 DRAIN FLASK PER PUMP -(I|17.29)-[∠]INTEGRATED -34.50--31.98 D & C 3.12 -MANUAL DRAIN T29.07 LIFTING EYES MANUAL-Date: 2/10/2014 A 40.05 Date: Drawn: JWO 1/2" FNPT CONDENSATE This drawing and the information contained thereon remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering. 8.38 -60.75 -68.00 DRAIN **-75.25** 4107 8549 33 GENERAL ASSEMBLY 15HP DX CLAW VSD VAC SPC 240G RECEIVER HOP230420 Sheet 1 of 1 DO NOT SCALE THIS DOCUMENT Form F-007 Rev. 01 8