



Approx. Mass $\pm 110$ (LBS)		Center of Gravity $\pm 2$ (inch)	
Type	Pack	Dim.	Pack
ZR37VSD	2915	W	19.8
ZR55VSD	2998	L	40.8
		H	27.8

All materials supplied are in compliance with the requirements of the List of Prohibited Substance




 Note:  
\* With full open door


02		See Note X	2011-05-20	
Ed	Position	Gewijzigd van Modified from	Datum Date	Ingev./Goedg. Int./Agreed

9820577300	02.10		Version Dwg 02.11	Blank wt.	0 Kg	Finl wt.	444.135 Kg	Designation	Sheet 1 / 1
Parent 3D model	Ed. Version 3D	Approved	Des checked.	Prod checked.	Approved.	Date	2008-01-22	9820577300	

5.  INDICATES CENTER OF GRAVITY.

1. ALLOW A MINIMUM OF (2) FEET ON ALL SIDES OF DRYER PACKAGE FOR PROPER VENTILATION AND MAINTENANCE.
2. BASEFRAMES ARE SHIPPED TOGETHER AND CAN BE UNBOLTED FOR PASSAGE THROUGH A STANDARD DOOR.
3. SHIPPING WIGHT: 2362 LBS.
4. SYSTEM CAPACITY: 318 CFM.






- NOTES:
- NEEDLE VALVE SETTING INSTRUCTIONS
    - ADJUST VALVE TO FULL OPEN.
    - CLOSE KNOB 2 FULL TURNS.
    - LOCK ADJUSTMENT KNOB WITH LOCK COLLAR.

Revised: BKW	Date: 6/30/2017	Drawn: BKW	Date: 4/18/2017
<b>BEACONMEDAES</b>	This drawing and the information contained therein remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering.		Scale: NA
Description: GEN ASM CD150 DRYER MODULE 318 CFM, Z-MED MEDICAL AIR SYSTEM		Drawing Number: 4107 8561 69	
DN: HOP170587		Rev: 01	
Sheet 2 of 2		DO NOT SCALE THIS DOCUMENT	





Revised:	Date:	Drawn: MDB	Date: 7/8/2021
BEACONMEDES	This drawing and the information contained therein remain the property of BeaconMedes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedes Engineering.		Scale: 1/10
Description: AIR RECEIVER TANK - 400 GALLON W/STAND ALONE SIGHT GLASS & DRAIN		Drawing Number: 4107 0082 32	Rev: 00
Sheet 1 of 1		DN: HOP211010	DO NOT SCALE THIS DOCUMENT