




CONFIDENTIAL:
This document is property of Atlas
copied, used for manufacturing or




NOTES:

● - CENTER OF GRAVITY

Tolerances, if not indicated, according to:						
ATLAS COPCO STANDARD CLASS						
Name		Dimension Drawing			GA7-15 VSD+ MED NFPA	
Material		Not Applicable			Confidentiality Class acc. to 1102 K	
Treatment		Not Applicable			Internal	
					INV	
		Scale 1 : 10		Family		Compare
		Drawn by INEmis		Blank nr.		Replaces
		Drawing format		Fin. wt.		Designation
STATUS		A2		Kg		Sheet 1 / 1
For Approval		Des checked.		Prod checked.		Approved.
				Date		9/26/2019
						9820695008

All materials supplied are in compliance with the requirements of the List of Prohibited Substances

CONFIDENTIAL:
This document is our property and shall not without our permission be altered, copied, used for manufacturing or communicated to any other person or company.




NOTES:
- "VESSEL" DEFINES AS THE SUPPLY INLET FLOW OF DIRECTION.
- "LINE" DEFINES AS THE OUTLET FLOW OF DIRECTION.



CONNECTION INFORMATION			
PART NUMBER	DESCRIPTION	SIDE 1	SIDE 2
Y1	INLET SOLENOID	X	
Y2	TOWER A SOLENOID VALVE	X	
Y3	TOWER B SOLENOID VALVE	X	
Y4	INLET SOLENOID		X
Y5	TOWER A SOLENOID VALVE		X
Y6	TOWER B SOLENOID VALVE		X
PT01	VESSEL PRESSURE TRANSDUCER		X
PT02	LINE PRESSURE TRANSDUCER		X
P1	PRESSURE SWITCH	X	
P2	PRESSURE SWITCH		X

CONNECTIONS		
TP No.	DESCRIPTION	CONNECTION SIZE
TP01	AFTER FILTER AIR OUTLET	G1"
TP02	PRE FILTER AIR INLET	G1"

SIDE 1 SIDE 2





Title: dMED 035 (11) NFPA Instrument air Dimension Drawing								
Material:						Third angle		
Finish:								
		REMOVE ALL BURR AND SHARP EDGES UNLESS OTHERWISE STATED		DIMENSIONS IN MILLIMETERS		A2	PGC CODE:	DRAWING OWNER GBNPBA
				TOLERANCES UNLESS OTHERWISE STATED			DRAWING SCALE: 1 : 10	
Status		DO NOT SCALE		NO DECIMAL PLACE: ± 0.25 ± 1.0 ONE DECIMAL PLACE: ± 0.1 ± 0.5 TWO DECIMAL PLACE: ± 0.05 ± 1.0		PART NUMBER:		SHEET: 1 OF 1
WorkInProgress		IF IN DOUBT ASK		ANGULAR: ± 0.5 ± 1.0 SURFACE FINISH:		2212021751		

1. ASME CODE CONSTRUCTION.
2. DESIGN PRESSURE 250 PSI @ 400°F.
3. NATIONAL BOARD REGISTERED.
4. ALL OPENINGS ARE NPT THREADED.
5. RECEIVER APPROXIMATELY 613 LBS.




Revised:	GTM	Date:	5/20/2020	Drawn:	CCC	Date:	2/18/2020
 BEACONMEDES		This drawing and the information contained thereon remain the property of BeaconMedes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedes Engineering.				Scale:	1:8
Description: 200 GAL ASM W/1.5" BYP & 1.5" CONNS, SIGHT GLASS & DRAIN, INST AIR APP		Drawing Number: 4107 8572 46					
		DN: HOP200305				Rev: 01	
		Sheet 1 of 1				DO NOT SCALE THIS DOCUMENT	

All materials supplied are in compliance with the requirements of the List of Prohibited Substances



Note:
- All dimensions are in imperial units

Tolerances, if not indicated, according to:									
ATLAS COPCO STANDARD CLASS									
Name DIMENS. DRWG						OSC 95	Confidentiality Class acc. to 1102 K Confidential 		
Material									
Treatment Not Applicable						INV			
		Scale 1 : 10		Family		A3	Compare	Drawing owner API	
		Drawn by AIR27282		Blank nr.					Replaces
		Version Drwg		Blank wt		Kg	Finl wt. 6.447 kg		Designation
STATUS		Des checked.		Prod checked.		Approved.		Date	9829529100-01
Pending								20/11/2018	

CONFIDENTIAL:

This document is our property and shall not without our permission be altered, copied, used for manufacturing or communicated to any other person or company.

1					9829529100-M1	
Ed	Posi-tion	Modified from	Date	Intr./Appd.	Parent 3D model	Ed . Version 3D