



Notes:

- 1. DISCHARGE MUFFLERS WILL BE SHIPPED LOOSE.
- 2. SYSTEM WEIGHT: 4,471 LBS.
- 3.  INDICATES LIFTING POINT.
- 4.  DENOTES CENTER OF GRAVITY.

SYSTEM CONNECTIONS	
ID	DESCRIPTION
I	INLET LOCATION
D	DISCHARGE LOCATION
C	CONDENSATE DRAIN LOCATION
T	TANK DRAIN LOCATION

MINIMUM CLEARANCE FOR  
MAINTENANCE & VENTILATION

Ø0.75 BOLTDOWNS  
(12 PLACES)

ELECTRICAL  
CONNECTION

RECEIVER BYPASS  
VALVE

200 GALLON  
RECEIVER

BIOHAZARD DRAIN  
FLASK PER PUMP

INLET  
3" FLANGE

DISCHARGE  
2" FLANGE  
(2 PLACES)

CONTROL CABINET

HEPA INTAKE FILTER  
PER PUMP

MANUAL DRAIN  
1/2" FNPT

INTEGRATED  
LIFTING EYES


MANUAL  
CONDENSATE  
DRAIN, TYP

Revised: SMB

Date: 6/24/2019

Drawn: MDB

Date: 2/4/2010

 BEACONMEDAES

This drawing and the information contained thereon remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering.

Description:  
GENERAL ASSEMBLY 7.5HP QX  
CLAW VAC SPC 200G RECEIVER  
STANDARD & O2 ASSURED

Part Number:  
4107 8513 84

DN: HOP210808

Rev: 02

Sheet 1 of 1

DO NOT SCALE THIS DOCUMENT