

Notes:

- DISCHARGE MUFFLERS WILL BE SHIPPED LOOSE.
- WEIGHT OF DUPLEX MODULE W/ PANEL: 2,312 LBS
WEIGHT OF DUPLEX MODULE W/O PANEL: 2,144 LBS
(INCLUDES FUTURE EXPANSION PUMP)
200G TANK WEIGHT: 658 LBS
- ☒ INDICATES LIFTING POINT.
- DENOTES CENTER OF GRAVITY.
(INCLUDES FUTURE EXPANSION PUMP)

SYSTEM CONNECTIONS	
ID	DESCRIPTION
I	INLET LOCATION
D	DISCHARGE LOCATION
C	CONDENSATE DRAIN LOCATION
T	TANK DRAIN LOCATION

ELECTRICAL CONNECTION

FUTURE EXPANSION PUMP

REMOVE INTAKE MANIFOLD BEFORE LIFTING (2 PLACES)

REMOVE JUNCTION BOX BEFORE LIFTING (LOCATED ON THIS MODULE ONLY)

MINIMUM CLEARANCE FOR MAINTENANCE & VENTILATION

Ø0.75 BOLTDOWN (8 PLACES)

OUTLET 4" FLANGE

RECEIVER BYPASS VALVE

HOSPITAL CONN 4" FLANGE

MANUAL DRAIN CONN 1/2" NPT (FEMALE)

0.69 X 1.06 SLOT ON 26.00 DIA B.C. EQUALLY SPACED (4 PLACES)

200 GAL VACUUM RECEIVER 4" FLANGED CONNECTIONS (SCALE 1:10)

CONTROL CABINET

INLET 3" FLANGE (2 PLACES)

HEPA INTAKE FILTER PER PUMP

DISCHARGE 3" FLANGE (2 PLACES)

JUNCTION BOX FOR CONTROL WIRING

INTAKE ISOLATION VALVE FOR FUTURE EXPANSION PUMP

BIOHAZARD DRAIN FLASK PER PUMP

INTEGRATED LIFTING EYES

MANUAL CONDENSATE DRAIN, TYP

Revised: APS	Date: 4/17/2023	Drawn: MDB	Date: 5/14/2012
BEACONMED-ES		This drawing and the information contained therein remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering.	
Description: GENERAL ASSEMBLY 10HP TX-X-QX CLAW VAC MOD 200G RECEIVER STANDARD AND O2 ASSURED		Part Number: 4107 8539 20	
DN: HOP230841		Rev: 03	
Sheet 1 of 1		DO NOT SCALE THIS DOCUMENT	