Notes:

1. TX MODULE WEIGHT: 2,732 LBS
   200G TANK WEIGHT: 638 LBS
2. ⊠ INDICATES LIFTING POINT.
3. ⊕ DENOTES CENTER OF GRAVITY.

SYSTEM CONNECTIONS	
ID	DESCRIPTION
I	INLET LOCATION
D	DISCHARGE LOCATION
C	CONDENSATE DRAIN LOCATION
T	TANK DRAIN LOCAITON



MINIMUM CLEARANCE FOR MAINTENANCE & VENTILATION



200 GAL VACUUM RECEIVER
3" FLANGED CONNECTIONS
(SCALE 1:10)







HEPA INTAKE FILTER PER PUMP

MANUAL CONDENSATE DRAIN

Revised: APS	Date: 3/3/2023	Drawn: USNKW	Date: 10/25/2010
BEACONMEDÆS	This drawing and the information contained thereon remain the property of BeaconMedaes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedaes Engineering.		Scale: D
Description: GENERAL ASSEMBLY 10HP TX LUBE VAC MOD 200G RECEIVER		Part Number: 4107 8522 19	
		DN: HOP230842	Rev: 03
		Sheet 1 of 1	DO NOT SCALE THIS DOCUMENT