



| INSTRUMENTATION DETAILS | | | |
|-------------------------|------------------|----------------|------------------|
| SENSOR | DESCRIPTION | SETTING | ACTION |
| S1 | HIGH DISCH TEMP | 10HP-465°F | SHUTDOWN |
| | | 15HP-465°F | |
| S2 | PRESS TRANSDUCER | 0-145PSI | OPERATION |
| S3 | BACKUP PRESS SW. | CUT IN-110PSI | BACKUP OPERATION |
| | | CUT OUT-135PSI | |
| S4 | DP SENSOR | +2' C | ALARM |

| | | | |
|--|------------------|--|-----------------|
| Revised: SMB | Date: 11/18/2019 | Drawn: JDY | Date: 1/27/2011 |
| | | This drawing and the information contained thereon remain the property of BeaconMedes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedes Engineering. | |
| Description: PNEUMATIC DIAGRAM QX SCROLL 10BAR SPC W/DRYER (10-15HP) | | Part Number: 4107 8526 50 | Scale: NTS |
| Dwg: HOP191299 | Rev: 03 | DO NOT SCALE THIS DOCUMENT Sheet 1 of 1 | |