



Notes:

- TX MODULE WEIGHT: 1,553 LBS  
120G TANK WEIGHT: 393 LBS
- ☒ INDICATES LIFTING POINT.
- DENOTES CENTER OF GRAVITY.

SYSTEM CONNECTIONS	
ID	DESCRIPTION
I	INLET LOCATION
D	DISCHARGE LOCATION
C	CONDENSATE DRAIN LOCATION
T	TANK DRAIN LOCATION





MINIMUM CLEARANCE FOR MAINTENANCE & VENTILATION



120 GAL VACUUM RECEIVER  
2" FLANGED CONNECTIONS  
(SCALE 1:10)



CONTROL CABINET  
INTEGRATED LIFTING EYES



HEPA INTAKE FILTER PER PUMP  
MANUAL CONDENSATE DRAIN

Revised: APS	Date: 3/3/2023	Drawn: USNKW	Date: 10/21/2010
		This drawing and the information contained therein remain the property of BeaconMedics and may not be used for other than the purpose for which it is issued without the expressed written permission from BeaconMedics Engineering.	
Description: GENERAL ASSEMBLY 5HP TX LUBE VAC MOD 120G RECEIVER		Part Number: 4107 8522 14 DN: HOP230842 Rev: 04	
Sheet 1 of 1		DO NOT SCALE THIS DOCUMENT	