



INSTRUMENTATION DETAILS			
SENSOR	DESCRIPTION	SETTING	ACTION
S1	HIGH INLET VACUUM	2" Hg	SHUTDOWN
S2	HIGH DISCH TEMP	495°F	SHUTDOWN
S3	PRESS TRANSDUCER	0-145PSI	OPERATION
S4	BACKUP PRESS SW.	CUT IN -80PSI CUT OUT -105PSI	BACKUP OPERATION
S5	DP SENSOR	+2° C	ALARM
S6	CO SENSOR	10PPM	ALARM

Revised: TAR	Date: 05/12/2021	Drawn: TAR	Date: 12/13/2019
		This drawing and the information contained therein remain the property of BeaconMedes and may not be used for other than the purpose for which it is loaned without the expressed written permission from BeaconMedes Engineering.	
Description: PNEUMATIC DIAGRAM PX SCROLL SPC W/DRYER (20HP)		Part Number: 4107 8567 07	
Dwg: HOP210917		Rev: 00	
Sheet: 1 of 1		DO NOT SCALE THIS DOCUMENT	